Electret Blank Measurement Log

Date	Initial Voltage	Date	Final Voltage	Difference	Electret Serial Number

Perform blank measurement for 5 percent of all radon measurements. Read initial voltage, transport device with devices being placed for measurement, return blank device to office. Read voltage upon return of measurement devices. Voltage drop should not exceed 2 volts per week. Notify QA officer for investigation if reading drops by more than 2 volts.	